

Reprocessed by



Sustainability Solutions

Instructions for Use Reprocessed HoverMatt®

Reprocessed Device for Single Patient Use

Caution: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

- NON-STERILE
- Exposed to Ethylene Oxide (EO) gas
- Not made with natural rubber latex

Explanation of Symbols

	Date of Reprocessing
	Product Code
	Lot Number
	Do Not Reuse
	See Instructions For Use
	Not made with natural rubber latex
	Non-Sterile
	Prescription Only

Reprocessed HoverMatt® Description

The Stryker Sustainability Solutions Reprocessed HoverMatt® Air Transfer Mattress is an inflatable mattress used to assist caregivers with lateral transfers, repositioning, and proning of patients. The Reprocessed HoverMatt® is inflated to cushion and cradle the patient while air simultaneously escapes from the holes on the underside, reducing the force needed to move the patient.

The Reprocessed HoverMatt® is intended for use in hospitals and long term or extended care facilities.

Indications for Use

The Reprocessed HoverMatt® is indicated for use with patients unable to assist in their own lateral transfer and for patients whose weight or girth poses a potential health risk for the caregivers responsible for repositioning or laterally transferring said patients.

Contraindications for Use

The Reprocessed HoverMatt® is contraindicated for patients with unstable thoracic, cervical or lumbar fractures that are deemed unstable, unless used in conjunction with a spinal board on top of the mattress (follow your state's protocol with respect to the use of spinal boards).

Warnings and Precautions

- Inspect device for damage prior to use.
- Use only compatible HoverTech Air Supplies for inflation of the Reprocessed HoverMatt®. Read and follow all instructions in the user manual for compatible HoverTech Air Supplies.
- Prior to inflating the mattress, ensure bed/stretcher rails are raised.
- Prior to inflating the mattress, ensure the patient is centered on the mattress and the mattress is centered on the support surface.
- Caregivers must verify that all caster brakes have been engaged prior to transfer.
- Additional caregivers are recommended when moving a patient over 750 lbs. / 340kg.
- The weight limit for the Reprocessed HoverMatt® is 1200 lbs. The weight limit for the Reprocessed HoverMatt® Half-Matt is 600 lbs.
- Never attempt to move a patient on an uninflated mattress.
- Route the power cord in a manner to ensure freedom from hazard.
- Ensure connection between Air Supply and mattress is secure before inflating.
- Avoid blocking the air intakes of the Air Supply.
- Never leave patient unattended on an inflated device.
- When transferring to a low air loss bed, set the bed mattress air flow to the highest level for a firm transfer surface.
- When using in an MRI environment, a 25 ft. specialty MRI hose is required.
- For safety, when using the Reprocessed HoverMatt® always use a minimum of two caregivers during patient transfer. When using the Reprocessed HoverMatt® Half-Matt, always use a minimum of three caregivers during patient transfer.
- Use this product only for its intended purpose as described in this manual.
- In the Operating Room – to prevent the patient from slipping always deflate the Reprocessed HoverMatt® and secure the patient and HoverMatt® to the OR table prior to moving the table into an angled position.
- Do not launder the single-patient use Reprocessed HoverMatt®.

Directions for Use

1. Patient should be in a horizontal position for transfer/repositioning on the appropriate width mattress [34" w (86cm), 39" w (99cm), or 50" (127cm)]. Regardless of ease of patient movement, always use a minimum of two caregivers for transfer with the Reprocessed HoverMatt® and a minimum of three caregivers for transfer with the Reprocessed HoverMatt® Half-Matt.
2. Place the transfer mattress underneath patient using a log-rolling technique and secure patient safety straps loosely. Whatever the patient is lying on to keep the hospital bed clean may also be placed on top of the Reprocessed HoverMatt®.
3. When using the Reprocessed HoverMatt® Link, loosely attach connecting straps to solid points on the bed frame. Before lateral transfers and positioning, disconnect the connecting straps from the bed frame and stow in corresponding storage pockets.
4. Use HoverTech Air Supplies compatible with associated transfer mattresses.
5. Plug the Air Supply power cord into an electrical outlet.
6. Insert the Air Supply hose nozzle into the mattress in one of the two entries at foot-end of the transfer mattress and snap in place.
7. Ensure that transfer surfaces are as close as possible and lock all wheels.
8. If possible, transfer from a higher surface to a lower surface.

9. Turn on Air Supply.
10. Grasp transfer handles and push the Reprocessed HoverMatt® at an angle, either head first or feet first. Once half-way across, opposite caregiver should grasp closest handles and pull to the desired location.
11. When using the Reprocessed HoverMatt® Half-Matt, ensure caregiver at foot-end guides patient's feet during the transfer.
12. Ensure that the patient is centered on the receiving equipment prior to deflation, especially if the width of the equipment receiving the patient is less than the width of the transfer mattress.
13. Turn off Air Supply to deflate the mattress and employ the bed/stretcher rails. Unfasten patient safety straps.

Using the Reprocessed HoverMatt® in the Operating Room

- **Option 1** - Place the HoverMatt® on the Pre-Op stretcher or bed prior to patient arrival. Have patient ambulate onto bed/stretcher or use a HoverMatt® to perform a lateral transfer. Once in the OR, ensure the OR table is secured and locked to the floor, then transfer the patient onto the OR table. Have a caregiver at head-end of OR table and ensure the patient is centered before deflating the HoverMatt®. Position the patient as needed for surgery. Tuck the edges of the HoverMatt® under the OR table pad, and ensure table rails are accessible. For supine surgeries, follow your facility's patient positioning protocol. After the case, release the edges of the HoverMatt® from under the OR table. Buckle patient safety straps loosely. Partially inflate the HoverMatt® using the ADJUSTABLE setting, have head-end caregiver ensure the patient is centered, then fully inflate using the appropriate high speed setting. Transfer the patient to the stretcher or bed.
- **Option 2** - Prior to patient arrival, place the HoverMatt® on the OR table and tuck the edges under the OR table pad. Ensure table rails are accessible. Transfer the patient onto the table, and proceed as described in Option 1.
- **Trendelenburg Position** - If Trendelenburg or Reverse Trendelenburg is required, an appropriate anti-slide device that secures to the frame of the OR table must be used. For Reverse Trendelenburg, a device that clamps to the OR table frame, such as a footplate, should be used. If the surgery also includes a tilt side-to-side (airplaning), the patient must be safely secured to accommodate this position prior to starting surgery.

Warranty

Reprocessed Products

Stryker warrants all reprocessed products, subject to the exceptions provided herein, to be free from defects in reprocessing and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for one use in accordance with the instructions for use of such product.

STRYKER SHALL NOT BE LAIBLE FOR ANY DAMAGES TO THE EXTENT CAUSED BY ANY DEFECT IN MATERIAL, WORKMANSHIP OR DESIGN BY THE ORIGINAL MANUFACTURER OF THE PRODUCT OR ANY ACT OR OMISSION OF THE ORIGINAL MANUFACTURER OF THE PRODUCT.

Products for which Stryker is the Original Manufacturer

Stryker warrants all products for which it is the original manufacturer, subject to the exceptions provided herein, to be free from defects in design, materials and workmanship and to substantially conform to the product specifications contained in the documentation provided by Stryker with the products for a period of one year from the date of purchase.

General Warranty Terms Applicable to All Products

TO THE FULLEST EXTENT PERMITTED BY LAW, THE EXPRESS WARRANTY SET FORTH HEREIN IS THE ONLY WARRANTY APPLICABLE TO THE PRODUCTS AND IS EXPRESSLY IN LIEU OF ANY OTHER WARRANTY BY STRYKER, EXPRESSED OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OR MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. IN NO EVENT WILL STRYKER'S LIABILITY ARISING IN CONNECTION WITH THE SALE OF THE PRODUCT (WHETHER UNDER THE THEORIES OF BREACH OF CONTRACT, TORT, MISREPRESENTATION, FRAUD, WARRANTY, NEGLIGENCE, STRICT LIABILITY OR ANY OTHER THEORY OF LAW) EXCEED THE PURCHASE PRICE, CURRENT MARKET VALUE OR RESIDUAL VALUE OF THE PRODUCTS, WHICHEVER IS LESS. STRYKER SHALL NOT BE LIABLE FOR INDIRECT, SPECIAL, INCIDENTAL, PUNITIVE OR CONSEQUENTIAL DAMAGES RESULTING FROM ANY BREACH OF WARRANTY OR UNDER ANY OTHER LEGAL THEORY.

This warranty shall apply only to the original end-user purchaser of products directly from Stryker or a Stryker authorized distributor. This warranty may not be transferred or assigned without the express written consent of Stryker.

This warranty does not apply to: (1) products that have been misused, neglected, modified, altered, adjusted, tampered with, improperly installed or refurbished; (2) products that have been repaired by any person other than Stryker personnel without the prior written consent of Stryker; (3) products that have been subjected to unusual stress or have not been maintained in accordance with the instructions in the user manual or as demonstrated by a Stryker representative; (4) products on which any original serial numbers or other identification marks have been removed or destroyed; or (5) products that have been repaired with any unauthorized or non-Stryker components.

If a valid warranty claim is received within thirty (30) days of the expiration of the applicable warranty period, Stryker will, in its sole discretion: (1) replace the product at no charge with a product that is at least functionally equivalent to the original product or (2) refund the purchase price of the product. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property. In any event, Stryker's liability for breach of warranty shall be limited to the replacement value of the defective or non-conforming part or component.

If Stryker determines in its reasonable discretion that the claimed defect or non-conformance in the product is excluded from warranty coverage as described hereunder, it will notify the customer of such determination and will provide an estimate of the cost of repair of the product. In such an event, any repair would be performed at Stryker's standard rates.

Products and product components repaired or replaced under this warranty continue to be warranted as described herein during the initial applicable warranty period or, if the initial warranty period has expired by the time the product is repaired or replaced, for thirty (30) days after delivery of the repaired or replaced product. When a product or component is replaced, the item provided in replacement will be the customer's property and the replaced item will be Stryker's property. If a refund is provided by Stryker, the product for which the refund is provided must be returned to Stryker and will become Stryker's property.

The OEM information listed on the label is provided as device ID prior to reprocessing and may contain the trademarks of unrelated third parties that do not sponsor this device.

This product and its packaging have been exposed to ethylene oxide gas (EtO). Even though the product then is processed in compliance with all applicable laws and regulations relating to EtO exposure, Proposition 65, a State of California voter initiative, requires the following notice:

Warning: This product and its packaging have been exposed to ethylene oxide. The packaging may expose you to ethylene oxide, a chemical known to the State of California to cause cancer or birth defects or other reproductive harm.

HoverMatt® is a registered trademark of HoverTech International.

HHM EL10094 Rev. B 09/2019